

**AUSTRALIAN CAPITAL TERRITORY****RADIATION ACT 1983****Section 73****NOTIFICATION OF DECISIONS**

Pursuant to Section 73 of the Australian Capital Territory Radiation Act 1983, the Radiation Council hereby gives notice of the particulars of the following decisions of Council made on 6<sup>th</sup> December 2000.

**GRANTING OF LICENCE (paragraph 73(1)(b))**

Licensee	Jennet Bramley
Address	4 Faulkner Place CHAPMAN ACT 2611
Authorised Activities	Use, cause or permit to use, irradiating apparatus.
Conditions	Diagnostic Radiology. You are required to enter the total radiation exposure time of each fluoroscopic procedure into the patient's record.

**REGISTRATION OF IRRADIATING APPARATUS (paragraph 73(1)(d))**

Owner	ACT X-ray Services
Description	Sirona model D3297 diagnostic x-ray machine (OPG/Ceph) with model 5968573D3200 tube S/Nos 50850, 26833 90kVp, 12mA
Location	Lakeview Square 21 Benjamin Way Belconnen
Conditions	Diagnostic radiography.

Owner	Calvary Public Hospital
Description	Philips BV 300 (Mobile II) x-ray machine with Philips 4202040/000380 tube, 105kVp S/Nos 150141005476, 953233
Location	Medical Imaging Department Level 2, Xavier Building Calvary Public Hospital, BRUCE
Conditions	Diagnostic radiography.

Owner	Calvary Public Hospital
Description	Philips Practix 2000 (Mobile) x-ray machine with Philips Emerald tube, S/Nos 4512-590-10160, AS 34089/RAD 4 125kVp 40mA
Location	Medical Imaging Department Level 2, Xavier Building Calvary Public Hospital, BRUCE
Conditions	Diagnostic radiography.

Owner	Calvary Public Hospital
Description	Philips Duo Diagnost (II) x-ray machine with Philips RO 17/50 tube S/Nos 0005052, 655136 655620 150kVp 750mA
Location	Medical Imaging Department Level 2, Xavier Building (Screening Rm) Calvary Public Hospital, BRUCE
Conditions	Diagnostic radiography.
Owner	Calvary Public Hospital
Description	Philips Practix 2000 x-ray machine with Philips Emerald tube S/Nos 4512-590-10161, AS 34089/RAD 4 125kVp 400mA
Location	Medical Imaging Department Level 2, Xavier Building (Room 2) Calvary Public Hospital, BRUCE
Conditions	Diagnostic radiography.
Owner	ACT X-ray Services Pty Ltd
Description	Toshiba CXXG-008A/1A (CT Scanner) x-ray machine with Varian CXB-750C tube S/Nos AO562194, 14490-OS 135kVp 400mA
Location	CT Room, X-ray Department John James Memorial Hospital DEAKIN
Conditions	Diagnostic radiography.
Owner	Calvary Public Hospital
Description	Siemens diagnostic x-ray CT Scanner model 4806118K1140 with Man Siemans tube model 4791005k1138 S/Nos 24183, 099180002 140kVp 500mA
Location	Medical Imaging Department Level 2, Xavier Building (CT) Calvary Public Hospital, BRUCE
Conditions	Diagnostic radiography.
Owner	Calvary Public Hospital
Description	Philips Optimus 50 x-ray machine model 989000002001 with model 980630071202 tube. S/Nos 0246, 95515G/955634 150kVp 500mA
Location	Medical Imaging Department Level 2, Xavier Building (Room 1) Calvary Public Hospital, BRUCE
Conditions	Diagnostic radiography.

**APPROVAL FOR ALTERATION OR MODIFICATION OF REGISTERED  
IRRADIATING APPARATUS (paragraph 73(1)(e))**

Owner	Fred Hayes
Description	Morita diagnostic dental (OPG) x-ray machine model Panex EC with Toshiba G1005 tube. S/Nos 6H005, 18121-6103 90kVp, 40mA
Location	Kippax Health Centre HOLT ACT
Conditions	Diagnostic dental radiography.
Owner	ACT X-ray Services Pty Ltd
Description	Toshiba CT Scanner model CXXC-005A with Varian CXB-350A tube. S/Nos A6572064, 27038-S6 130kVp, 300mA
Location	Valley Diagnostic Centre, CT Room Gratton Court ERINDALE
Conditions	Diagnostic radiography.

**Review by the ACT Administrative Appeals Tribunal** - Any Persons whose interests are adversely affected by this decision may apply to the Tribunal to have the decision reviewed.

**Reasons** - If you wish to obtain a statement of reasons to explain why the decision was made you should write within 28 days of this notice to the ACT Radiation Council, GPO Box 825, Canberra 2601.

**Location of the ACT Administrative Appeals Tribunal** - The Tribunal is located on the 4th Floor, Canberra House, 40 Marcus Clarke Street, Canberra City.

Postal Address: GPO Box 9955, Canberra 2601. Telephone: 6243 4611. Facsimile: 6247 0962. Document Exchange: DX 5727.

**Powers of the ACT Administrative Appeals Tribunal** - The Tribunal is an independent body. The Tribunal can agree with, change or reject the original decision, substitute its own decision or send the matter back to the decision maker for reconsideration in accordance with Tribunal recommendations.

**How to apply to the ACT Administrative Appeals Tribunal** - Simply write within 28 days explaining the details to the decision and the reasons for asking for a review.

**Cost** - To lodge an application there is a fee. You may apply to have the fee waived on the grounds of hardship. No fee is payable if you are receiving legal assistance.

**Access to documents** - You may apply for access to any documents relevant to this decision under the ACT *Freedom of Information Act 1989*. For more information contact the Freedom of Information Officer, Department of Health and Community Care, on phone 62051340.

Dated this 6<sup>th</sup> day of December 2000.

**Joseph Lising**  
**Chairperson**  
**Radiation Council**