

RADIATION ACT 1983

Section 73

NOTIFICATION OF DECISIONS

Pursuant to Section 73 of the Australian Capital Territory Radiation Act 1983, the Radiation Council hereby gives notice of the particulars of the following decisions of Council, made on 4 August 1993.

GRANTING OF LICENCE (paragraph 73(1)(b))

Licensee: John Dominic Connors

Address: Canberra Imaging

PO Box 3035

Belconnen ACT 2617

Authorized Activities: Sell, Purchase, own or have in possession, use,
cause or permit to use irradiating apparatus

Conditions: Diagnostic radiology

Licensee: Peter Gordon Stuart Walmsley

Address: 6 Wenholtz Street

Farrer ACT 2607

Authorized Activities: Own or have in possession, use, cause or permit to
use irradiating apparatus

Conditions: Diagnostic dental radiography

REGISTRATION OF IRRADIATING APPARATUS (paragraph 73(1)(d))

Owner : Belconnen Imaging Unit Trust

Description : Norland x-ray bone densitometer
Model D051, S/No. 7014
100kVp, 1mA

Location : 16-18 Purdue Street
Belconnen

Conditions : Bone density analysis

Owner : CSIRO Division of Soils

Description : Micromeritics Sedigraph 5100 particle size
x-ray analysis system
S/Nos A132, 849 13.6kVp, 3mA

Location : CSIRO Division of Soils
Building 301
Black Mountain Laboratories

Conditions : X-ray analysis

Owner : ACT Health, Dental Services

Description : Philips densomat dental x-ray machine with
Oralix 65S tube
S/Nos 8923223, 9000623 65kVp, 7.5mA

Location : School Dental Therapy Clinic
Mt Rogers Community School
Spence Campus (Baddely Cres)

Conditions : Diagnostic dental examinations

Owner :ACT Health, Dental Services
Description :Trophy dental x-ray machine model CCX with
Model 708 tube
S/Nos 291085, G7502
Location :Adult Dental Clinic, 1st Floor
Surgery 2, ACT Health Building
1 Moore Street, Canberra
Conditions :Diagnostic dental examinations

Owner :ACT Health, Dental Services
Description :Trophy dental x-ray machine model CCX with
Model 708 tube
S/Nos 291085, G7502
Location :Adult Dental Clinic, 1st Floor
Surgery 4, ACT Health Building
1 Moore Street, Canberra
Conditions :Diagnostic dental examinations

GRANTING OF DISPOSAL PERMIT (Paragraph 73(1)(g))


Permit Holder: Dr M J Whitten
Licensed Premises: Clunies Ross Street, Acton
CSIRO Div. of Entomology

RADIONUCLIDE contained in the material for disposal.	FORM of the material.	Annual QUANTITY of material for which approval is granted.	Disposal METHOD approved.
Phosphorus-32	Liquid and solid	500MBq	Burial, W.B.L.C.
Carbon-14	Waste in a range	37MBq	Burial, W.B.L.C.
Hydrogen-3	of chemical	37MBq	Burial, W.B.L.C.
Sulphur-35	forms	37MBq	Burial, W.B.L.C.
Iodine-125		450kBq	Burial, W.B.L.C.
Phosphorus-32	Water Soluble	185MBq	Sewer
Iodine-125	Water Soluble	450MBq	Sewer, J.C.S.M.R.

The Radiation Council hereby gives notice that subject to the provisions of the Administrative Appeals Tribunal Act 1975, a person whose interests are adversely affected by a decision of the Radiation Council is entitled to apply to the Administrative Appeals Tribunal for a review of the decision.

Dated this 4th day of August 1993.

R D BROCK
Chair
Radiation Council


4/8/93