

AUSTRALIAN CAPITAL TERRITORY

RADIATION ACT 1983

Section 73

NOTIFICATION OF DECISIONS

Pursuant to Section 73 of the Australian Capital Territory Radiation Act 1983, the Radiation Council hereby gives notice of the particulars of the following decisions of Council, made on 13th March 2001.

REGISTRATION OF IRRADIATING APPARATUS (paragraph 73(1)(d))

| | |
|-------------|---|
| Owner | Peripar Pty Ltd |
| Description | Vincennes Trophy 1Rix70 dental x-ray machine with Trophy 9430 Type 708 tube. |
| Location | S/Nos CZ843068, 1100437 70kVp, 7mA Unit 15, (Rm 4) McKay Gardens Professional Centre Turner |
| Conditions | Diagnostic dental examinations. |
| Owner | Peripar Pty Ltd |
| Description | Gendex model 765DC dental x-ray machine with GX-70-10DC tube. S/Nos 344, 10-1483179DP 65kVp, 7mA |
| Location | Unit 15, (Rm 3) McKay Gardens Professional Centre Turner |
| Conditions | Diagnostic dental examinations. |
| Owner | Dr Naren Chellappah |
| Description | Gendex 765DC dental x-ray machine with Gendex GX70-10 tube, S/Nos 1056, 10-1476884 DP. 65 kVp, 7 mA |
| Location | Unit 2 24 Mahony Circuit Weston |
| Conditions | Diagnostic dental examinations. |
| Owner | CSR Emoleum |
| Description | CPN moisture/density gauge model MC-3 Portaprobe. Calibrated 26-5-99/11-8-99 Cs137, 370MBq, Am241/Be, 1.87GBq. S/No MC-300205481 |
| Location | CSR Emoleum 109-111 High Street Queanbeyan NSW |
| Conditions | Soil density/moisture analysis. |

| | |
|-------------|---|
| Owner | Dr Sharon Mathai |
| Description | Gendex model 765DC dental x-ray machine with model GX 70-10 DC tube S/Nos 111479918, 368. 65kVp, 7.5mA |
| Location | Room 2 1/32 Furneaux Street Manuka |
| Conditions | Diagnostic dental examinations. |

**APPROVAL FOR ALTERATION OR MODIFICATION OF REGISTERED
IRRADIATING APPARATUS (paragraph 73(1)(e))**

| | |
|-------------|---|
| Owner | Dr Robert Kwee Keong Low |
| Description | Dental x-ray unit Castellini Victory X 65 COX 6698 CEI S/Nos 60144,1070355. 65kVp, 8mA |
| Location | Suite 7 1st Floor Dickson Chambers, Dickson |
| Conditions | Diagnostic dental radiography. |

| | |
|-------------|---|
| Owner | QANTAS Airways Limited |
| Description | Baggage security x-ray system EG & G Astrophysics model 01-0421, S/No. 50058 System Ten E-Scan. 160kVp 1mA |
| Location | Qantas Departure Gate Canberra Airport |
| Conditions | Baggage inspections. |

| | |
|-------------|---|
| Owner | Dr Robert Kwee Keong Low |
| Description | Planmeca Prostyle Intra dental x-ray machine S/Nos EIC952253, 952575, 3894. 70kVp, 8mA |
| Location | Suite 7, First Floor Dickson Chambers Dickson |
| Conditions | Diagnostic dental examinations. |

**VARIATION OF CONDITIONS SPECIFIED IN A LICENCE
(paragraph 73(1)(c))**

| | |
|------------------------|---|
| Licensee | Dr Iain E T Stewart |
| Address | National Capital Diagnostic Imaging: (1) Corinna Chambers, Corinna Street, Woden. (2) Derwent House, University Avenue, Canberra City. |
| Authorised Activities: | Purchase, own or have in possession, use, cause or permit to use, unsealed radioactive materials, irradiating apparatus. |
| Conditions | * Diagnostic radiology. You are required to enter the total radiation exposure time of each fluoroscopic procedure into the patient's record. * Sealed sources: Co-57, Cs-137. * Nuclear medicine procedures involving diagnostic and therapeutic uses of: Mo-99, Tc-99m, Tl-201, Ga-67, I-131. |

GRANTING OF LICENCE (paragraph 73(1)(b))

| | |
|------------------------|--|
| Licensee | Robert Moran |
| Address | 19 Clem Hill Street GORDON ACT 2906 |
| Authorised Activities: | Sell, manufacture, use, irradiating apparatus. |
| Conditions | The sale, installation, servicing and testing of medical x-ray equipment. |
| Licensee | Thuraisamy Ravichander |
| Address | Department of Radiation Oncology The Canberra Hospital GARRAN |
| Authorised Activities: | Purchase, own or have in possession, use, cause or permit to use, unsealed radioactive materials, irradiating apparatus. |
| Conditions | As Hospital Physicist - Supervision of medical uses of:- * Therapeutic and diagnostic irradiating apparatus. * Sealed and unsealed radioactive materials: I-131, Sr-89, Cs-137, Sr-90, Tc-99m, Ga-67, Tl-201, Ir-192, Mo-99, I-125, Cr-51, Co-57, Se-75, Y-90, P-32, Co-60, In-111. |
| Licensee | Dr Tran Vinh Lam |
| Address | Suite 1001 AMP Building Hobart Place CANBERRA ACT 2601 |
| Authorised Activities: | Use, cause or permit to use, irradiating apparatus. |
| Conditions | Diagnostic dental radiography. |
| Licensee | Andrew Ho |
| Address | 1/1 Whittle Street HUGHES ACT 2605 |
| Authorised Activities: | Own or have in possession, use, cause or permit to use, irradiating apparatus. |
| Conditions | Diagnostic dental radiography. |

Review by the ACT Administrative Appeals Tribunal - Any Persons whose interests are adversely affected by this decision may apply to the Tribunal to have the decision reviewed.

Reasons - If you wish to obtain a statement of reasons to explain why the decision was made you should write within 28 days of this notice to the ACT Radiation Council, GPO Box 825, Canberra 2601.

Location of the ACT Administrative Appeals Tribunal - The Tribunal is located on the 4th Floor, Canberra House, 40 Marcus Clarke Street, Canberra City.

Postal Address: GPO Box 9955, Canberra 2601. Telephone: 6243 4611. Facsimile: 6247 0962. Document Exchange: DX 5727.

Powers of the ACT Administrative Appeals Tribunal - The Tribunal is an independent body. The Tribunal can agree with, change or reject the original decision, substitute its own decision or send the matter back to the decision maker for reconsideration in accordance with Tribunal recommendations.

How to apply to the ACT Administrative Appeals Tribunal - Simply write within 28 days explaining the details to the decision and the reasons for asking for a review.

Cost - To lodge an application there is a fee. You may apply to have the fee waived on the grounds of hardship. No fee is payable if you are receiving legal assistance.

Access to documents - You may apply for access to any documents relevant to this decision under the ACT *Freedom of Information Act 1989*. For more information contact the Freedom of Information Officer, Department of Health and Community Care, on phone 6205 1340.

Dated this 13th day of March 2001

Anthony Agostino
Deputy Chairperson
Radiation Council